

Open Access

Abstract

Published 03/06/2024

Copyright

© Copyright 2024

Selvaraj et al. This is an open access abstract distributed under the terms of the Creative Commons Attribution License CC-BY 4.0., which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Distributed under

Creative Commons CC-BY 4.0

Thoracic Spinal Cord FLASH SBRT Treatment Using a Single-Energy Proton Pristine Bragg Peak Delivery Technique

Balaji Selvaraj ¹, Minglei Kang ², Xingyi Zhao ¹, Alexander Bookbinder ¹, Haibo Lin ³, J. Isabelle Choi ⁴, Arpit M. Chhabra ⁴, Shaakir Hasan ⁴, Charles B. Simone ⁴, Yoshiya Yamada ⁵

1. Physics, New York Proton Center, New York, USA 2. Medical Physics, New York Proton Center, New York, USA 3. Medical Physics, New York Proton Center, New York, NY, USA 4. Radiation Oncology, New York Proton Center, New York, USA 5. Radiation Oncology, Memorial Sloan Kettering Cancer Center, New York, USA

Corresponding author: Balaji Selvaraj, bselvaraj@nyproton.com

Categories: Medical Physics, Radiation Oncology

Keywords: thoracic spinal cord flash sbrt

How to cite this abstract

Selvaraj B, Kang M, Zhao X, et al. (March 06, 2024) Thoracic Spinal Cord FLASH SBRT Treatment Using a Single-Energy Proton Pristine Bragg Peak Delivery Technique. Cureus 16(3): a1186

Abstract

Objectives:

Treating a complex doughnut-shaped target with the spinal cord in the middle is challenging, even when using advanced proton pencil beam scanning (PBS). FLASH-RT may help to reduce the risk of complications for spine stereotactic body radiation therapy (SBRT), but the feasibility of using a single-energy proton pristine Bragg peak delivery technique to deliver conformal FLASH-RT to thoracic spine patients has not been studied. This study aims to examine the dosimetric and dose rate performance of this novel approach, termed Bragg peak FLASH (BP-FLASH), for thoracic spine SBRT.

Methods:

Thoracic spine patients previously treated using conventional multiple-energy intensity-modulated proton therapy (CONV-IMPT) to 40 Gy in 5 fractions were re-optimized based on an in-house developed FLASH platform using the single-energy proton Bragg peak technique. BP-FLASH plans were designed with a similar beam arrangement to CONV-IMPT. The dose distributions and dose metrics were compared between the two techniques. The dose rate of the BP-FLASH plans was computed based on clinically proton machine parameters, including the max proton beam current, PBS spot delivery and scanning time. A quasi-mean dose rate calculation method averaged dose rate (ADR) was used to evaluate voxel-based 40Gy/s dose rate coverage (V40Gy/s) for OARs FLASH-sparing effect assessment.

Results:

The dose coverage of BP-FLASH plans were normalized similarly to CONV-IMPT plans. All the dose metrics for major OARs, including spinal cord, esophagus, lung, and heart, were comparable between BP-FLASH and CONV-IMPT. The CTV Dmax was slightly higher in FLASH plans (BP-FLASH: 106.1% vs. CONV-IMPT: 102.2%) due to the absence of energy modulation for BP-FLASH compared to CONV-IMPT. The dose rate volume histograms (DRVHs) indicated that more than 50% of the volume of the OARs can reach a dose rate of at least 40Gy/s with no dose threshold applied, and 90% of the volume can reach a dose rate greater than 40Gy/s with a 2Gy dose threshold applied.

Conclusion(s):

BP-FLASH plans can deliver highly conformal doses to the target while sparing the spinal cord, comparable to multiple-energy CONV-IMPT. Treating a complex doughnut-shaped target with a critical OAR inside using Bragg peak FLASH is feasible.