

Open Access

Abstract

Published 03/05/2025

Copyright

© Copyright 2025

Ram et al. This is an open access abstract distributed under the terms of the Creative Commons Attribution License CC-BY 4.0., which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Distributed under

Creative Commons CC-BY 4.0

Evaluation of High-Fidelity Mode for Semi-Automated Multi-Met, Single-Isocenter Stereotactic Radiosurgery Planning Using the Ethos 2.0 Planning System

Udbhav Ram ¹, Joel A. Pogue ², Richard Popple ³, Natalie Viscariello ², Dennis Stanley ⁴, John Fiveash ⁴, Carlos Cardenas ⁴

¹. Radiation Oncology, McMaster University, Hamilton, Ontario, CAN ². Radiation Oncology, University of Alabama at Birmingham, Birmingham, USA ³. Radiation Oncology, University of Alabama at Birmingham School of Medicine, Birmingham, USA ⁴. Department of Radiation Oncology, Heersink School of Medicine, University of Alabama at Birmingham, Birmingham, USA

Corresponding author: Udbhav Ram, ramu@mcmaster.ca

Categories: Medical Physics, Radiation Oncology

Keywords: semi-automated multi-met, single-isocenter, stereotactic radiosurgery

How to cite this abstract

Ram U, Pogue J A, Popple R, et al. (March 05, 2025) Evaluation of High-Fidelity Mode for Semi-Automated Multi-Met, Single-Isocenter Stereotactic Radiosurgery Planning Using the Ethos 2.0 Planning System. Cureus 17(3): a1420

Abstract

Objectives:

The recently released Ethos 2.0 treatment planning system (TPS) includes updates for the semi-automatic generation of stereotactic plans. In this study, we evaluate the impact of the new 'high-fidelity mode' feature on the quality of CNS stereotactic radiosurgery (SRS) plans. The high-fidelity mode within the Ethos treatment planning system (TPS) employs a high-resolution dose grid (1.25 mm), allows for higher dose concentration within the target, and aims to achieve sharper dose fall-off while minimizing plan complexity. An attractive feature of the Ethos treatment platform is that it can accommodate and treat patients weighing up to 500 lbs, making single-isocenter SRS a viable option for patients with higher body weight who may not be eligible for treatment on other machines due to weight limitations.

Methods:

Fifty brain metastasis patients previously treated with linac-based SRS were anonymized and imported into an Ethos 2.0 emulator. Fifteen of these cases were used to optimize Ethos 2.0 planning templates, which were then applied without further adjustment to the remaining 35 patients. All plans were prescribed 30 Gy in 5 fractions, with a 2 mm margin from GTV to PTV. Planning templates with and without rings were evaluated in high fidelity mode and compared to standard templates with and without rings. Control rings are commonly used in our clinical practice to enhance dose shaping around the target, helping to limit exposure to surrounding healthy tissues by creating sharper dose gradients. Their inclusion or exclusion in the planning process was assessed to determine their impact on plan quality and dose distribution. Key clinical metrics analyzed included per-target PTV coverage (V100%), whole-brain dose outside the PTV (Brain-PTV V24Gy), Paddick Gradient Index (GI), and RTOG Conformity Index (CI). A Wilcoxon paired statistical analysis was performed to quantify differences across the templates.

Results:

The tuning and testing cohorts included cases with a range of 1 to 13 and 1 to 17 targets, respectively. Enabling both high-fidelity mode and control rings resulted in superior average metrics: PTV V100% of 99.4%, Brain-PTV V24Gy of 24 cc, CI of 1.17, and GI of 4.0. In contrast, when both high-fidelity mode and control rings were disabled, the average values were 98.2% for PTV V100%, 29.21 cc for Brain-PTV V24Gy, CI of 1.34, and GI of 3.97. Notably, the addition of control rings in combination with high-fidelity mode significantly improved Brain-PTV V24Gy and CI scores (p < 0.01).

Conclusion(s):

The findings of this study indicate that enabling high-fidelity mode and utilizing control rings on the Ethos 2.0 treatment planning system can improve the quality of single-isocenter, multi-target SRS plans compared to those created without high-fidelity mode. This approach may offer a viable treatment option for patients with higher body weight who may not be eligible for conventional linac-based SRS due to equipment weight limits, thus expanding access to precise stereotactic treatments. Further studies are warranted to validate these findings and confirm their efficacy and safety in routine clinical practice.

