



Open Access Abstract Published 03/05/2025

Copyright

© Copyright 2025

Zhu et al. This is an open access abstract distributed under the terms of the Creative Commons Attribution License CC-BY 4.0., which permits unrestricted use, distribution and reproduction in any medium, provided the original author and source are credited.

Distributed under Creative Commons CC-BY 4.0

Rapid VMAT LATTICE Therapy Planning Using Deep Learning Predicted Synthetic CT from Diagnostic CT Scans

Libing Zhu ¹, Nathan Y. Yu ², Ali Ammar ³, Quan Chen ³, Yi Rong ³

1. Radiation Oncology, Mayo Clinic, Jacksonville, USA 2. Radiation Oncology, Santa Clara Valley Medical Center, San Jose, USA 3. Radiation Oncology, Mayo Clinic, Phoenix, USA

Corresponding author: Libing Zhu, zhu.libing@mayo.edu

Categories: Medical Physics, Radiation Oncology Keywords: ct scans, vmat lattice therapy

How to cite this abstract

Zhu L, Yu N Y, Ammar A, et al. (March 05, 2025) Rapid VMAT LATTICE Therapy Planning Using Deep Learning Predicted Synthetic CT from Diagnostic CT Scans. Cureus 17(3): a1487

Abstract

Objectives:

Volumetric Modulated Arc Therapy (VMAT) lattice therapy is utilized to treat bulky malignant tumors because of its biological effectiveness and local tumor control. This study aims to develop an expedited treatment process by utilizing synthetic CT (sCT) images generated through deep learning for VMAT lattice therapy.

Methods:

Two deep learning models, based on a 3D-UNet architecture, were trained to predict sCT for the thoracic and abdominal regions. VMAT lattice therapy plans were created on the 15 sCT cases and recalculated on the pCT. Clinical dose-volume histogram (DVH) metrics were used to assess dosimetric differences between the sCT and pCT, including D0.03cc for organ-at-risks (OARs), peak-to-valley dose ratio (PVDR), D10%, D50%, D90%, maximum, minimum and mean dose of gross tumor volume (GTV), and D50% of all spheres. Statistical significance between two CTs was determined using the Wilcoxon signed-rank test.

Results:

No statistically significant differences were found in the DVH metrics for the organs-at-risk in the thoracic and abdominal regions (p > 0.05). Among the GTV DVH metrics, only the minimum dose and D10% showed statistical significance, while PVDR, D50%, D90%, mean dose, and maximum dose did not. The mean absolute deviation between sCT and pCT was 0.32 Gy for GTV D10%, 0.18 Gy for D50%, 0.13 Gy for D90%, 0.41 Gy for PVDR, 0.43 Gy for the maximum dose, 0.23 Gy for the mean dose, 0.17 Gy for the minimum dose, and 0.40 Gy for all spheres' D50%.

Conclusion(s):

The deep learning-generated sCT images were highly similar to pCT images, and the dosimetric comparison between sCT and pCT showed no significant differences in most key DVH metrics. This demonstrates the potential of using sCT to streamline treatment planning and accelerate VMAT lattice therapy delivery.