

Received 10/14/2022 **Review began** 11/15/2022 **Review ended** 01/03/2023 **Published** 01/14/2023

© Copyright 2023

Mandhalkar et al. This is an open access article distributed under the terms of the Creative Commons Attribution License CC-BY 4.0., which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

DOI: 10.7759/cureus.33779

Application of Nanomaterials in Restorative Dentistry

Rutvik [Mandhalkar](https://www.cureus.com/users/428833-rutvik-n-mandhalkar) 1 , [Priyanka](https://www.cureus.com/users/176968-priyanka-paul) Paul 2 , Amit [Reche](https://www.cureus.com/users/370956-amit-reche) 2

1. Dentistry, Sharad Pawar Dental College and Hospital, Datta Meghe Institute of Medical Sciences, Deemed to be University, Wardha, IND 2. Public Health Dentistry, Sharad Pawar Dental College and Hospital, Datta Meghe Institute of Medical Sciences, Deemed to be University, Wardha, IND

Corresponding author: Rutvik Mandhalkar, rutvikmandhalkar23@gmail.com

Abstract

Dental composite resins are widely popular restoratives, as, when using these tools to restore the tooth, only the infected and affected carious structures are removed. This allows the patient to retain a greater quantity of their natural tooth structure than they would have using conventional principles of cavity preparation. Nanomaterials are a new concept concerning the manipulation of materials on an atomic or molecular level. However, on a nanoscale, the chemical, biological, and physical properties of an atom vary compared to the properties of its naturally occurring compound form. The main idea of shifting focus to the inclusion of nanomaterials is to aid in the detection, treatment, and prevention of the recurrence of a pathology (secondary caries). The primary aim of using nanomaterials in composites is to augment their strength, wear resistance, and microhardness. This usage also reduces polymerization shrinkage. Nanomaterials are capable of enhancing mechanical properties, life, and bond strength between dentin and restoration. This review aims to highlight different research studies and experiments that have been conducted on the use of nanomaterials in restorative dentistry in order to understand the versatility of these materials and their viability in practice.

Categories: Public Health, Epidemiology/Public Health, Dentistry **Keywords:** nanoparticles, caries, anti-bacterial, composite resin, restoration

Introduction And Background

Tooth restoration is one of the main procedures in dental treatment. It is concerned with the reestablishment of the normal structure of the teeth such that the treated or restored teeth mimic the nature and function of the original non-affected or non-infected teeth. The classic cavity preparation principles were established by the late Dr. G.V Black. His principles of tooth preparation aimed to guide an operator to prepare a cavity such that when the restoration is placed, it assumes the function of the tooth perfectly and stays inside the cavity for a long time and mimics the natural behavior and function of the tooth. The cavity prepared on the principles of Dr. Black was, for the most part of the endodontic history, filled using amalgam. Amalgam is an alloy of silver and mercury in a 3:7 ratio [\[1\]](javascript:void(0)). It was continuously improved by adding filler particles. Amalgam has been the principal restorative material used for most of endodontic history, but with the evolution of technology and knowledge, it has now been widely substituted with dental composite resins, a better, more versatile, aesthetic, and comparatively stronger alternative.

Dental composite resins are widely popular because they allowed operators to abandon the conventional principles of tooth preparation of Dr. Black for a far more conservative approach wherein only the infected and affected carious structures are removed; thus, in many cases, with the use of these resins, patients could retain a greater quantity of their natural tooth structures as compared to the use of conventional principles.

Dental composite resins are commonly composed of dimethacrylate monomers (Bis-GMA, TEGMA, and UDMA), along with multi-functional filler materials, which help enhance elastic modulus, increase strength and wear resistance, and decrease the polymerization shrinkage of the restoration [\[1\]](javascript:void(0)). These filler particles can be classified as macro fillers, micro fillers, micro-hybrid fillers, and nanofillers. Much like amalgam was improved using copper, composites were improved using various filler materials, one of which is nanofillers. This article aims to review the improvements made by modern nano-materials to dental composite resins compared to other filler materials.

Why nano-materials?

Nano-materials are used to manipulate materials on an atomic or molecular level. However, on a nanoscale, the chemical, biological, and physical properties of an atom vary compared to its naturally occurring compound forms [\[1\]](javascript:void(0)). Human tooth structures include enamel that is made of 96% hydroxyapatite of size ranging from 10nm to 200 nm [\[2\].](javascript:void(0)) Intertubular dentin is 2-5nm in thickness and 60nm in length. Enamel and dentin are connected by collagen fibrils that stretch over 20-75 nm in length [\[3\]](javascript:void(0)). These characteristics provide a solid ground for the basic need for research applications of nano-materials in dentistry. The use of nano-materials in dentistry is known as nano dentistry $[4,5]$. The main idea of shifting focus to the

How to cite this article

Mandhalkar R, Paul P, Reche A (January 14, 2023) Application of Nanomaterials in Restorative Dentistry. Cureus 15(1): e33779. DOI 10.7759/cureus.33779

application of nano-materials is to aid in the detection, treatment, and prevention of the recurrence of pathology (secondary caries). The dawn of the digital age and increased patient needs and demands for aesthetics and functionality have posed a challenge to the field of engineered nano-materials [\[6\]](javascript:void(0)). Nanocomposites are biomaterials in which at least one phase shows dimensions in the nanometer range. The size of nano-materials ranges from 5nm to 100nm. The incorporation of such filler particles leads to color alteration and increases the flexural strength of the composite resin restoration [\[7\]](javascript:void(0)). Currently, a plethora of nano-materials is being experimentally studied to determine their ability to enhance products of wideranging disciplines. In dentistry, particularly restorative dentistry, nano-materials are used to manufacture nanocomposites [\[8\]](javascript:void(0)), glass ionomer cement, and endodontic sealer. The main purpose of using nanomaterials in composites is to augment their strength, wear-resistance, and microhardness, as well as to reduce the polymerization shrinkage [\[9\]](javascript:void(0)); however, polymerization shrinkage varies depending on the chemical structure and manufacturing process of the composite [\[10\]](javascript:void(0)). However, research shows that the optimum wt% of nanofillers should not be exceeded, as no improvement in mechanical properties can be achieved further [\[6\]](javascript:void(0)). Nano-materials are capable of enhancing the mechanical properties, life, and bond strength between dentin and restoration [\[11\]](javascript:void(0)).

The synthesis of nano-materials is mainly done in two ways: in the top-down method, a bulk of the material is reduced to nanometric measurements using various methods, and in the bottoms-up method, individual atoms are clumped together [\[11\]](javascript:void(0)). The classification of nano-materials is illustrated in Figure *[1](javascript:void(0))* [\[11\]](javascript:void(0))

Review

Research methodology

Primarily, articles and case studies were searched and obtained from the databases PubMed and Google Scholar. The inclusion criterion was studies that discussed the use of nanoparticles or the scope of nanoparticles in restorative dentistry only; further, the selected studies should also have discussed their results. The PRISMA flow diagram of the included studies is shown in Figure *[2](javascript:void(0))*.

FIGURE 2: PRISMA flow diagram.

PRISMA: Preferred Reporting Items for Systemic Reviews and Meta-Analyses

Discussion

The interaction of various nano-materials with resin restorations has exhibited the ability to improve various aspects of such restorations.

Xiao et al. [\[12\]](javascript:void(0)) developed a bioactive multifunctional composite (BMC) using nanoparticles of NACP, 2 methacryloyloxyethyl phosphorylcholine (MPC), dimethylaminohexadecyl methacrylate (DMAHDM), and silver (AgNP) and investigated the effect of poly(amido amine) (PAMAM), which was added to the BMC. The focus of the study was on root caries and demonstrated excellent root dentin remineralization. This study successfully developed a bioactive multi-functional composite for Class V restoration. Along with remineralization, the developed BMC was protein-repellent and had anti-bacterial capabilities [\[12\]](javascript:void(0)). It concluded that BMC + PAMAM protects tooth root structures and is promising for Class I, Class II, and other types of tooth cavity restorations [\[12\]](javascript:void(0)).

Ghahremani et al. [\[13\]](javascript:void(0)) evaluated a color-modified heat cure resin that was enhanced with titanium oxide (TiO₂) nanoparticles for tensile and impact strength. In this in vitro study, the researchers added TiO₂ nanoparticles to a triplex hot heat-cure resin and blended them together. Subsequently, it was found that the reinforced group's strength is much greater, specifically 7 MPa more than the strength of the control

group; this difference was statistically significant ($p = 0.001$). Thus, the study concluded that the addition of TiO² by 1% wt to color-modified acrylic resin increased tensile and impact strength; however, this study did not consider any side effects this approach may have caused on the restorative resin and its restorative ability [\[13\].](javascript:void(0))

In a comparative study performed by Meena et al. [\[14\],](javascript:void(0)) the researchers added marble powder to composite resin and compared its performance to composite added with nano alumina. They performed dynamic mechanical analysis and thermos-gravimetric analysis to compare the role of both components to the composite resin and its performance as a restoration. It was observed that marble dust-filled composites absorbed less water and had a lesser diffusion coefficient. Thermal stability of the nano-alumina-filled composite was also found to be lacking by 20% compared to composite filled with marble dust. The results drawn out from their analysis indicated that marble dust is a superior filler material as it showed better physical and mechanical properties compared to nano alumina; moreover, marble dust is a more economic and cheaper alternative to nano alumina [\[14\]](javascript:void(0)). However, further study is needed to see if restorations with marble dust-filled composites could have a long-standing effect on tooth biology.

Wang et al. [\[15\]](javascript:void(0)) incorporated wrinkled mesoporous silica (WMS) with unimodal and bimodal fillers. The resin matrix this was mixed with was BisGmA/TEGDMA based. Bimodal WMS fillers comprising either WMS-Si90 or WMS-Si190 were used to create resin composites, with all of them permitting 60 wt% filler loading and surpassing the loading limits of unimodal WMS (35 wt%). This study found that bimodal fillermixed WMS produced better mechanical properties compared to its unimodal filler counterpart [\[15\]](javascript:void(0)).

In the study by Al-Mosawi [\[16\]](javascript:void(0)), the researcher used ZnO nanoparticles, incorporated them into a composite resin, and evaluated their potential as an anti-bacterial agent for the oral cavity. He made his observations on agar at different concentrations such as 5%, 7%, and 10%; among all isolated bacterial species that demonstrated activity, the most vulnerable bacterial isolates were *S. mutans* and pseudomonas, respectively. These observations indicated that ZnO/NPs can be effective in secondary caries prevention [\[16\]](javascript:void(0)).

Ai et al. [\[17\]](javascript:void(0)) sought to discover a functional one-dimensional nanofibrous filler for composite resins with which effective reinforcing and excellent antibacterial activity can be achieved. By using calcium oleate as a precursor through the hydrothermal technique, hydroxyapatite (HA) nanowires were synthesized. These nanowires were coated in polydopamine (PDA) by soaking them in an aqueous dopamine solution. These were then treated with Ag nanoparticles. These AgNP-doped HA nanowires were mixed with Bis-GMA [50/50 w/w] and prepared through thermo-curing. These modified HA-PDA-Ag wires demonstrated great adherence with the Bis-GMA resin matrix and effectively reinforced the resin matrix. The resultant composite displayed high bactericidal activity without any display of cytotoxicity and served as an ideal nanofiller [\[17\].](javascript:void(0))

Paiva et al. [\[18\]](javascript:void(0)) aimed to develop polyacid formulations through photo-reducing Ag nanoparticles in a polyacrylate solution of standard glass ionomer cement (GIC) in a single step, retaining antibacterial activity and evaluating the handling and mechanical qualities of the Ag-added GIC as compared to a conventional GIC. The modified restoration was tested against *S. mutans*. This study found that the AgNPadded GIC had excellent antibacterial activity. The restoration acted on the principle of diffusion, indicating Ag ion dissolution (oxidative) from the cement matrix; this led to the conclusion that silver nanoparticlesadded GIC can stop caries and prevent biofilm growth on their surface [\[18\]](javascript:void(0)).

Stewart [\[19\]](javascript:void(0)) aimed to make the restoration itself antibacterial, which would make it capable of defending itself against the possibility of secondary caries. He incorporated octenidine dihydrochloride (OCT) in drugeluting mesoporous silica nanoparticles (DMSNs). These were fabricated using the bottoms-up synthesizing method. OCT is stated to be highly biocompatible [\[19\],](javascript:void(0)) and no microbial resistance has been discovered yet [\[20\]](javascript:void(0)). Using silica nanoparticles, the researchers developed a dental resin adhesive that was localized, that is, confined to the restored surface of the tooth, but long-time drug-releasing, thereby minimizing the systemic exposure. Their analysis, thus, proved that long-term prevention of secondary caries can be achieved through the restoration itself [\[21\]](javascript:void(0)).

Yue et al. [\[22\]](javascript:void(0)) developed a self-healing adhesive with anti-microbial and re-mineralizing abilities and evaluated the effect of involving microcapsules such as DMAHDM and the nanoparticles of NACP for the first time. Self-healing microcapsules were synthesized with polyurea-formaldehyde (PUF) shells that are composed of 10% DMAHDM and 20% NACP. The single-edge V-notched beam method was used to calculate fracture toughness, $\rm K_{ic}$, and crack-healing efficiency. Yue et al. [\[22\]](javascript:void(0)) achieved three results with the newly developed adhesive resin: autonomous crack-healing, anti-microbial capabilities, and remineralization function through the use of calcium phosphate nanoparticles. The fabricated adhesive also reduced the colony-forming units of microcosm biofilms by a magnitude of four grades compared to the conventional control. The unique strategy of employing triple agents (self-healing microcapsules + DMAHDM + NACP) to mend cracks and suppress caries is predicted to be suitable for a variety of dental adhesives, cement, sealants, and composites [\[22\].](javascript:void(0))

Cao et al. [\[23\]](javascript:void(0)) developed a nano Ag-filled resin. By using AgBr/BHPVP nanocomposites, any adverse effect on flexural strength and modulus was averted. Meanwhile, the Vicker's hardness of the resin discs was

greatly increased. There was a sustained Ag+ ion release without any impact from the anaerobic environment. The developed resin was suitable to be used to combat anaerobic cariogenic bacteria. It has a particularly potent antibacterial effect on *S. mutans*, which is achieved through the continuous release of Ag+ ions. The study concluded that the optimal concentration of AgBr/BHPVP should be 1.0 wt% in Bis-GMA/TEGDMA [\[23\]](javascript:void(0)).

The antibacterial effect of ZnO nanoparticles is higher than that of AgNPs with respect to harmful actions against *S. mutans* [\[24-26\]](javascript:void(0)). Additionally, nanodiamonds also show promising results. It was found that 1wt% quaternized copolymer functionalized nanodiamond-reinforced resin composites were capable of inhibiting biofilm formation without causing any harm to the tooth structure [\[27\].](javascript:void(0)) Not only composites, nanomaterials also have shown an excellent integration with GIC. GIC is inherently capable of releasing fluoride ions, which combat the secondary recurrence of caries [\[18\].](javascript:void(0)) Copper-doped GIC shows superior antibacterial properties and a reduced rate of collagen degradation. Renné et al. [\[28\]](javascript:void(0)) showed that the addition of nanohydroxyapatite powder to GIC enhanced its fluoride ion-releasing capability. Another study concluded that hydroxyapatite nanoparticles are suitable for improving the mechanical and bactericidal properties of GIC [\[29\]](javascript:void(0)).

Xie et al. [\[30\]](javascript:void(0)) aimed to develop an adhesive that has three advantageous properties: calcium phosphate ion recharging, protein repellent, and anti-bacterial function. MPC and DMAHDM were combined to create an NACP-rechargeable adhesive with protein-repellent and anti-bacterial properties to combat biofilms and caries. The developed bioactive adhesive had strong protein-repellent properties and significantly reduced bacterial adhesion. With these three advantages, the adhesive could protect tooth structures as well as reduce biofilm and caries progression [\[30\]](javascript:void(0)).

All these studies show that the addition of nano-materials to composite resins for restoration is advantageous to restorative dentistry (Table *[1](javascript:void(0))*).

effect.

TABLE 1: Literature reviewed in the study.

BMC: bioactive multifunctional dental composite

NP: nanoparticle

NACP: nanoparticles of amorphous calcium phosphate

DMAHDM: dimethylaminohexadecyl methacrylate

MPC: 2-methacryloyloxyethyl phosphorylcholine

NP: nanoparticle

TiO₂: titanium dioxide

ZnO: zinc oxide

HA: hydroxyapatite

AgNP: silver nanoparticles

GIC: glass ionomer cement

Ag+: silver ion

OCT: octenidine dihydrochloride

DSPs: dexamethasone sodium phosphate

NACP: amorphous calcium phosphate

AgBr: silver bromide

BHPVP: cationic polymer nanocomposite

Conclusions

The developments of different products using nanoparticles in the articles reviewed in this study proved that the demand and functions of nano-materials in restorative dentistry and dental practices are vast as a whole. The addition of various nanoparticles not only augmented the strength and life of restorations but also resisted secondary caries incidence by attributing antibacterial properties to the restorations. Besides, resin adhesive showed potential beyond assistance in restoration, functioning as a long-term local drug delivery tissue interface. Other than silver and titanium, nanoparticles have shown excellent biocompatibility with tooth; however, the level of integration with composite resin varied, although favorable results were obtained. While the addition of nanoparticles is certainly advantageous, the process of commercially manufacturing restorations infused with nanoparticles scale has not been thoroughly researched. An in-depth investigation needs to be conducted on the potential of the commercial manufacturing of nanoparticle-infused restorations, their competitive pricing on the market, and their availability to the general public while ensuring that the price of restoration as a medical service is not significantly impacted. Nanoparticles showed promise in various aspects of life; as such, as a new technology, nanoparticles have been researched extensively for their application in everyday life. This article reviewed the published studies on the use of nanoparticles in restorative dentistry to highlight the significance of this up-and-coming technology to the wider audience. While more clinical trials are needed to truly make this technology commercial, nano-materials still show the potential to emerge as a newgeneration restorative material in the field of dentistry.

Additional Information

Disclosures

Conflicts of interest: In compliance with the ICMJE uniform disclosure form, all authors declare the following: **Payment/services info:** All authors have declared that no financial support was received from

any organization for the submitted work. **Financial relationships:** All authors have declared that they have no financial relationships at present or within the previous three years with any organizations that might have an interest in the submitted work. **Other relationships:** All authors have declared that there are no other relationships or activities that could appear to have influenced the submitted work.

References

- 1. Roco MC: [Nanotechnology:](https://dx.doi.org/10.1016/s0958-1669(03)00068-5) convergence with modern biology and medicine . Curr Opin Biotechnol. 2003, 14:337-46. [10.1016/s0958-1669\(03\)00068-5](https://dx.doi.org/10.1016/s0958-1669(03)00068-5)
- 2. Deyhle H, Oliver B, Buser S, et al.: [Bio-inspired](https://dx.doi.org/10.1117/12.827437) dental fillings. Society of Photo-Optical Instrumentation Engineers (SPIE) Conference Series. 2009, 7401[:10.1117/12.827437](https://dx.doi.org/10.1117/12.827437)
- 3. Goldberg M, Kulkarni AB, Young M, Boskey A: Dentin: structure, composition and [mineralization](https://dx.doi.org/10.2741/e281). Front Biosci (Elite Ed). 2011, 3:711-35. [10.2741/e281](https://dx.doi.org/10.2741/e281)
- 4. Kumar PS, Kumar S, Savadi RC, John J: [Nanodentistry:](https://dx.doi.org/10.1007/s13191-011-0062-0) a paradigm shift-from fiction to reality . J Indian Prosthodont Soc. 2011, 11:1-6. [10.1007/s13191-011-0062-0](https://dx.doi.org/10.1007/s13191-011-0062-0)
- 5. Gaiser S, Deyhle H, Bunk O, White SN, Müller B: Understanding nano-anatomy of healthy and carious human teeth: a prerequisite for nanodentistry. Biointerphases. 2012, 7:4. [10.1007/s13758-011-0004-8](https://dx.doi.org/10.1007/s13758-011-0004-8)
- 6. Besinis A, De Peralta T, Tredwin CJ, Handy RD: Review of nanomaterials in dentistry: interactions with the oral [microenvironment,](https://dx.doi.org/10.1021/nn505015e) clinical applications, hazards, and benefits. ACS Nano. 2015, 9:2255-89. [10.1021/nn505015e](https://dx.doi.org/10.1021/nn505015e)
- 7. Ateyah N.: Mechanical behavior of water-aged nano-filled hybrid composite [restoratives](https://dx.doi.org/10.1016/j.ksujds.2012.11.004) . King Saud Univ J Dent Sci. 2013, 4:21-5. [10.1016/j.ksujds.2012.11.004](https://dx.doi.org/10.1016/j.ksujds.2012.11.004)
- 8. Mitra SB, Oxman JD, Falsafi A, Ton TT: Fluoride release and recharge behavior of a [nano-filled](https://pubmed.ncbi.nlm.nih.gov/22263336) resinmodified glass ionomer compared with that of other fluoride releasing materials. Am J Dent. 2011, 24:372-8.
- 9. Khurshid Z, Zafar M, Qasim S, Shahab S, Naseem M, AbuReqaiba A: Advances in nanotechnology for restorative dentistry. Materials (Basel). 2015, 8:717-31. [10.3390/ma8020717](https://dx.doi.org/10.3390/ma8020717)
- 10. Jain AK, Deepti D, Tavane PN, Singh A, Gupta PK, Gupta A, Sonkusre S.: Evaluation of [microleakage](https://www.semanticscholar.org/paper/Evaluation-of-Microleakage-of-Recent-Nano-hybrid-in-Jain-Deepti/3c8a5bb9cbe4d35d941d5e96bd19da4a6b8514c9) of recent nano-hybrid composites in class V restorations: an in vitro study. Int. J. Adv. Health Sci. 2015, 2:8-12.
- 11. Foong LK, Foroughi MM, Mirhosseini AF, et al.: Applications of [nano-materials](https://dx.doi.org/10.1039/d0ra00762e) in diverse dentistry regimes . RSC Adv. 2020, 10:15430-60. [10.1039/d0ra00762e](https://dx.doi.org/10.1039/d0ra00762e)
- 12. Xiao S, Liang K, Weir MD, et al.: Combining bioactive multifunctional dental composite with PAMAM for root dentin remineralization. Materials (Basel). 2017, 10:89. [10.3390/ma10010089](https://dx.doi.org/10.3390/ma10010089)
- 13. Ghahremani L, Shirkavand S, Akbari F, Sabzikari N: Tensile strength and impact strength of color modified acrylic resin reinforced with titanium dioxide [nanoparticles.](https://dx.doi.org/10.4317/jced.53620) J Clin Exp Dent. 2017, 9:e661-5. [10.4317/jced.53620](https://dx.doi.org/10.4317/jced.53620)
- 14. Meena A, Mali HS, Patnaik A, Kumar SR.: Effect of adding nanoalumina and marble dust powder on the physical, mechanical, and [thermo-mechanical](https://dx.doi.org/10.1002/pc.24131) characterization of dental composite. Polymer Composites. 2016, 39:E321-31. [10.1002/pc.24131](https://dx.doi.org/10.1002/pc.24131)
- 15. Wang R, Habib E, Zhu XX: Evaluation of the filler packing structures in dental resin composites: from theory to practice. Dent Mater. 2018, 34:1014-23. [10.1016/j.dental.2018.03.022](https://dx.doi.org/10.1016/j.dental.2018.03.022)
- 16. Al-Eisa A, Al-Mosawi M, Gupta R, D'Souza TM, Sharma PN.: Attitudes of general practitioners and pediatricians to management of acute [pyelonephritis](https://dx.doi.org/10.1159/000026128) in children. Med Princ Pract. 2000, 9:119-24. [10.1159/000026128](https://dx.doi.org/10.1159/000026128)
- 17. Ai M, Du Z, Zhu S, Geng H, Zhang X, Cai Q, Yang X: Composite resin reinforced with silver nanoparticlesladen [hydroxyapatite](https://dx.doi.org/10.1016/j.dental.2016.09.038) nanowires for dental application. Dent Mater. 2017, 33:12-22. [10.1016/j.dental.2016.09.038](https://dx.doi.org/10.1016/j.dental.2016.09.038)
- 18. Paiva L, Fidalgo TK, da Costa LP, et al.: Antibacterial properties and compressive strength of new one-step preparation silver nanoparticles in glass ionomer cements [\(NanoAg-GIC\).](https://dx.doi.org/10.1016/j.jdent.2017.12.003) J Dent. 2018, 69:102-9. [10.1016/j.jdent.2017.12.003](https://dx.doi.org/10.1016/j.jdent.2017.12.003)
- 19. Hübner NO, Siebert J, Kramer A: Octenidine dihydrochloride, a modern antiseptic for skin, mucous membranes and wounds. Skin Pharmacol Physiol. 2010, 23:244-58. [10.1159/000314699](https://dx.doi.org/10.1159/000314699)
- 20. Al-Doori Z, [Goroncy-Bermes](https://dx.doi.org/10.1093/jac/dkm092) P, Gemmell CG, Morrison D: Low-level exposure of MRSA to octenidine dihydrochloride does not select for resistance. J Antimicrob Chemother. 2007, 59:1280-1. [10.1093/jac/dkm092](https://dx.doi.org/10.1093/jac/dkm092)
- 21. Stewart CA, Hong JH, Hatton BD, Finer Y: Responsive antimicrobial dental adhesive based on drug-silica coassembled particles. Acta Biomater. 2018, 76:283-94. [10.1016/j.actbio.2018.06.032](https://dx.doi.org/10.1016/j.actbio.2018.06.032)
- 22. Yue S, Wu J, Zhang Q, et al.: Novel dental adhesive resin with crack self-healing, antimicrobial and remineralization properties. J Dent. 2018, 75:48-57. [10.1016/j.jdent.2018.05.009](https://dx.doi.org/10.1016/j.jdent.2018.05.009)
- 23. Cao W, Zhang Y, Wang X, et al.: Development of a novel resin-based dental material with dual biocidal modes and sustained release of Ag(+) ions based on photocurable core-shell AgBr/cationic polymer nanocomposites. J Mater Sci Mater Med. 2017, 28:103. [10.1007/s10856-017-5918-3](https://dx.doi.org/10.1007/s10856-017-5918-3)
- 24. Kasraei S, Sami L, Hendi S, Alikhani MY, Rezaei-Soufi L, Khamverdi Z: Antibacterial properties of composite resins incorporating silver and zinc oxide nanoparticles on [Streptococcus](https://dx.doi.org/10.5395/rde.2014.39.2.109) mutans and Lactobacillus. Restor Dent Endod. 2014, 39:109-14. [10.5395/rde.2014.39.2.109](https://dx.doi.org/10.5395/rde.2014.39.2.109)
- 25. Andrade V, Martínez A, Rojas N, Bello-Toledo H, Flores P, [Sánchez-Sanhueza](https://dx.doi.org/10.1016/j.prosdent.2017.09.015) G, Catalán A: Antibacterial activity against Streptococcus mutans and diametrical tensile strength of an interim cement modified with zinc oxide nanoparticles and terpenes: An in vitro study. J Prosthet Dent. 2018, 119:862.e1-7. [10.1016/j.prosdent.2017.09.015](https://dx.doi.org/10.1016/j.prosdent.2017.09.015)
- 26. Dutra-Correa M, Leite AA, de Cara SP, et al.: Antibacterial effects and cytotoxicity of an adhesive containing low concentration of silver nanoparticles. J Dent. 2018, 77:66-71. [10.1016/j.jdent.2018.07.010](https://dx.doi.org/10.1016/j.jdent.2018.07.010)
- 27. Cao W, Zhang Y, Wang X, et al.: Novel resin-based dental material with anti-biofilm activity and improved mechanical property by incorporating hydrophilic cationic copolymer [functionalized](https://dx.doi.org/10.1007/s10856-018-6172-z) nanodiamond. J Mater Sci Mater Med. 2018, 29:162. [10.1007/s10856-018-6172-z](https://dx.doi.org/10.1007/s10856-018-6172-z)
- 28. Renné WG, Lindner A, Mennito AS, et al.: Antibacterial properties of copper [iodide-doped](https://dx.doi.org/10.1007/s00784-016-1799-y) glass ionomer-

based materials and effect of copper iodide nanoparticles on collagen degradation. Clin Oral Investig. 2017, 21:369-79. [10.1007/s00784-016-1799-y](https://dx.doi.org/10.1007/s00784-016-1799-y)

- 29. Alatawi RA, Elsayed NH, Mohamed WS.: Influence of hydroxyapatite nanoparticles on the properties of glass ionomer cement. J Mater Res Technol.. 2019, 8:344-9. [10.1016/j.jmrt.2018.01.010](https://dx.doi.org/10.1016/j.jmrt.2018.01.010)
- 30. Xie X, Wang L, Xing D, et al.: Novel dental adhesive with triple benefits of calcium phosphate recharge, protein-repellent and antibacterial functions. Dent Mater. 2017, 33:553-63. [10.1016/j.dental.2017.03.002](https://dx.doi.org/10.1016/j.dental.2017.03.002)