

# The Origin and History of the N-Localizer for Stereotactic Neurosurgery

Russell A. Brown<sup>1</sup>, James A. Nelson<sup>2</sup>

1. Principal Engineer, A9.com, Palo Alto, USA 2. Radiology (Emeritus), University of Washington

✉ **Corresponding author:** Russell A. Brown, russ.brown@yahoo.com

Disclosures can be found in Additional Information at the end of the article

---

## Abstract

Nearly four decades after the invention of the N-localizer, its origin and history remain misunderstood. Some are unaware that a third-year medical student invented this technology. The following *conspectus* accurately chronicles the origin and early history of the N-localizer and corrects some misconceptions related to both.

---

**Categories:** Medical Physics, Radiation Oncology, Neurosurgery

**Keywords:** stereotactic neurosurgery, stereotactic radiosurgery, image guidance, image-guided, computed tomography, magnetic resonance imaging, positron emission tomography (pet), n-localizer, medical imaging, brain imaging

## Introduction And Background

The N-localizer has become an important tool for image-guided stereotactic neurosurgery and radiosurgery. The N-localizer produces two circles and one ellipse in tomographic images that are obtained via computed tomography (CT), magnetic resonance imaging (MRI) or positron emission tomography (PET). The relative spacing between the ellipse and the two circles precisely determines the position of the tomographic section with respect to the N-localizer (Figure 1) [1-2].

Received 08/18/2015

Review began 08/24/2015

Review ended 09/09/2015


Published 09/14/2015

© Copyright 2015

Brown et al. This is an open access article distributed under the terms of the Creative Commons Attribution License CC-BY 3.0., which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

### How to cite this article


Brown R A, Nelson J A (September 14, 2015) The Origin and History of the N-Localizer for Stereotactic Neurosurgery. Cureus 7(9): e323. DOI 10.7759/cureus.323



**FIGURE 1: N-Localizer and Its Interaction with the Tomographic Section**

(a) Side view of the N-localizer. The tomographic section intersects the N-localizer at two vertical rods and one diagonal rod. (b) Tomographic image. The intersection of the tomographic section with the N-localizer produces two circles and one ellipse. The relative spacing between the centers of the ellipse and the two circles varies according to the height at which the tomographic section intersects the diagonal rod. Measuring this spacing permits calculation of the position of the tomographic section with respect to the N-localizer [2].

Russell A. Brown invented the N-localizer in May 1978 when he was a third-year medical student and during a research elective under the supervision of James A. Nelson at the University of Utah [3]. In August 1978, Brown designed and built the first CT-compatible stereotactic frame in order to test the concept of the N-localizer (Figure 2). This stereotactic frame was presented at a joint meeting of the Western Neurological Society and the American Academy of Neurological Surgery held in Los Angeles, California in October 1978 [1] and at the INSERM Symposium on Stereotactic Irradiations held in Paris, France in July 1979 [4].



## FIGURE 2: The First CT-Compatible Stereotactic Frame

Brown designed and built this stereotactic frame in August 1978 in order to test the concept of the N-localizer [1]. Three N-localizers are attached to this frame and are merged end-to-end such that only seven rods are present. Because three points determine a unique plane in three-dimensional space, the locations of the centers of the three ellipses that are produced in a tomographic image by three N-localizers precisely determine the position of the tomographic section with respect to the stereotactic frame [2].

Beginning in 1979, six different stereotactic frames incorporated the N-localizer: the Brown-Roberts-Wells (BRW) frame [5], the Pfizer frame that was used at the University of Pittsburgh [6], the Kelly-Goerss frame [7], the modified Reichert-Mundinger frame that was used at Duke University [8], the Leksell frame [9], and the Cosman-Roberts-Wells (CRW) frame [10]. Subsequently, the N-localizer achieved widespread use in image-guided stereotactic neurosurgery and radiosurgery [11-40]. The simplicity and accuracy of the N-localizer render it an important tool for modern neurosurgery [3].

## Review

During the 37 years since the invention of the N-localizer, some misconceptions have arisen concerning its origin and history in relation to subsequent developments in image-guided stereotactic surgery.

The first misconception is that the Pfizer frame, which incorporated the N-localizer, was constructed and initially used in 1978. Kondziolka and Lunsford of the University of Pittsburgh assert this misconception, together with their failure to discuss the relevant literature, in their claim [41], "At our center, the first CT compatible stereotactic head frame, in collaboration with industry, was constructed in 1978 and utilized in 13 patients [6,42]. [...] During this interval, the newly redesigned Leksell CT compatible stereotactic head frame [43] was used for dedicated

brain biopsies under the direction of its inventor, Lars Leksell. Several groups were working on devices to allow accurate CT based stereotactic surgery [44]."

The above assertion presents an erroneous chronology. The Pfizer frame was neither the first CT-compatible stereotactic frame (Figure 2) nor was it constructed and initially used in 1978. Instead, it was constructed and initially used in 1979, as per Lunsford, *et al.*, who recount, "In 1979, our first efforts in image-guided stereotactic surgery attempted to adapt an early-generation Leksell frame. The metallic artifacts precluded adequate computerized tomography (CT) imaging, and we subsequently developed a CT-compatible stereotactic device (Pfizer frame [...]) [45,6] which was used in an initial series of 15 patients beginning in 1979" [26]. This statement is corroborated by Lunsford, Niranjana, Kassam, Khan, Amin and Kondziolka, who state, "During the interval of 1979 to 1980, 13 stereotactic procedures were performed in a diagnostic scanner at our hospital" [46]. These two statements confirm that the Pfizer frame was constructed and initially used in 1979, not in 1978.

The two above statements of Lunsford, *et al.* are corroborated by Perry, Rosenbaum, Lunsford, Swink and Zorub, who state that the Pfizer "stereotactic frame was made after attempts to modify the Leksell frame [...] proved difficult" [6]. Further corroboration is provided by a letter from Perry to Lunsford, Rosenbaum and Zorub [47] and a letter from Pfizer Medical Systems, Inc. to its patent attorney [48]. These letters verify that as of January 15, 1979, Perry, Rosenbaum, Lunsford and Zorub had not yet attempted any surgery using the modified early-generation Leksell frame. Hence, the Pfizer frame, which was constructed after efforts to adapt the early-generation Leksell frame had failed, was constructed no earlier than 1979.


The above assertion of Kondziolka and Lunsford disregards the fact that the CT-guidance technologies of the Leksell frame and the Pfizer frame were derivative. For both the Leksell and the Pfizer CT-compatible stereotactic frames, the inclusion of vertical and diagonal elements originated from Brown's prior invention and description of the N-localizer. This fact is established by the articles that introduced the Leksell [43] and Pfizer [6] frames. Both articles cited one [1] of Brown's original articles that had introduced the N-localizer one year earlier [1-2]. Although Lunsford (with and without Kondziolka) had previously cited [6,11,40,45-46,49] either of Brown's original articles [1-2], these coauthors cited neither of his original articles in their above assertion [41]. Instead, they cited a later article by Roberts and Brown [44] that was published contemporaneously with the first articles from the University of Pittsburgh [6,42] and one year after Brown's original articles had introduced the N-localizer [1-2].

The second misconception is that investigators from Pfizer and the University of Pittsburgh invented the N-localizer. This misconception is asserted by Lunsford, Niranjana, Kassam, Khan, Amin and Kondziolka, who claim [46], "During the subsequent years of training, the senior author had an opportunity to work with an innovative neuroradiologist, Arthur Rosenbaum, M.D., and an engineer, John Perry, Ph.D., who then headed the imaging division of Pfizer Medical Instruments. Together, we developed an image-guided stereotactic system using the now well-known N-localizer technology. This elegant solution was proposed by Perry *et al.* [6] and Rosenbaum *et al.* [42] independently and virtually simultaneously as publications from Brown [2] and Roberts and Brown [44] of Utah."

In the above assertion, the intended antecedent of "elegant solution" could be either "image-guided stereotactic system" or "N-localizer technology." Perry, *et al.* did propose the Pfizer image-guided stereotactic system [6] several months after Brown, *et al.* had proposed the Brown-Roberts-Wells (BRW) image-guided stereotactic system [5]. However, the historical record shows that none of the above-mentioned individuals, with the exception of Brown, invented the N-localizer. Instead, Perry adopted the N-localizer after Brown had disclosed it to him. Documents that corroborate these facts have remained preserved in the archives of the

U.S. Patent and Trademark Office for the past 30 years. The following discussion, which is based on those archives, recounts Perry's research related to image-guided stereotactic surgery and reveals the events that led to his adoption of the N-localizer.

Prior to the invention of the N-localizer, several coauthors had reported a method for estimating the position of a tomographic section with respect to patient anatomy [50-51]. That method involved a plate into which were milled vertical slots whose tops lay along a diagonal line (Figure 3).



**FIGURE 3: Slotted Plate and Its Interaction with the Tomographic Section**

(a) Side view of the slotted plate. The tomographic section intersects the plate into which are milled vertical slots of different lengths. The tops of the slots lie along a diagonal line. (b) Tomographic image. The intersection of the tomographic section with the slotted plate produces a variable number of notches. The number of notches depends on the height at which the tomographic section intersects the plate. Counting the number of notches permits estimation of the position of the tomographic section with respect to the slotted plate.

Documents from the archives of the U.S. Patent and Trademark Office indicate that as of

January 15, 1979, Perry, Rosenbaum, Lunsford and Zorub had attached three slotted plates to a Leksell frame [47-48]. In principle, three slotted plates could enable the calculation of the position of a tomographic section with respect to a stereotactic frame, similar to the manner in which three N-localizers enable this calculation (Figure 2).

In practice, however, the slotted plate was susceptible to error as a result of the discrete or quantized nature of the slots. Perry observed that it was necessary to manually count carefully the numerous notches that were visible in the tomographic image because any miscount would give rise to errors in the subsequent calculation of the position of the tomographic section with respect to the stereotactic frame [47]. Moreover, the partial volume effect [52-53], which derives from the several-millimeter thickness of the tomographic section, impeded accurate counting of the notches because any slot that passed partially into but not entirely through the tomographic section would produce an only faintly visible notch. For these reasons, the slotted plate was vulnerable to human error and hence was unsuitable for clinical use. The N-localizer avoids these quantization problems and the attendant possibility of computational errors by virtue of the continuous nature of the N-localizer's rods.

Perry's earliest report of the slotted plate, and indeed the earliest record of his involvement with image-guided stereotactic surgery, was in his letter dated January 15, 1979, and addressed to his collaborators, Lunsford, Rosenbaum and Zorub at the University of Pittsburgh [47]. Perry's letter describes three slotted plates attached to a stereotactic frame and provides instructions for using computer software in conjunction with those slotted plates to calculate the position of a tomographic section with respect to that stereotactic frame. Well before the date of Perry's letter, Brown had already invented the N-localizer [54], built the first CT-compatible stereotactic frame [55], and presented his results to the Western Neurological Society and the American Academy of Neurological Surgery [1].

On January 25, 1979, Brown spoke by phone with one of Perry's coworkers at Pfizer Medical Systems, Inc. and learned that Perry's research involved image-guided stereotactic surgery [56]. The following day, another of Perry's coworkers at Pfizer Medical Systems, Inc. sent to its patent attorney a letter that included a photo of a Leksell frame to which three slotted plates were attached and a photo of a CT scan image of the Leksell frame and slotted plates [48].

A few days following his conversation with Perry's coworker, Brown spoke by phone with Perry and disclosed the N-localizer to him [57]. Prior to this discussion with Brown, Perry had been unaware of the concept of the N-localizer. Perry may have apprised Rosenbaum of some aspects of this discussion with Brown. Nelson affirms that, during a conversation with Rosenbaum concerning the N-localizer, Rosenbaum revealed his awareness of Brown's prior discussion with Perry [57].

Several months following his discussion with Perry, Brown was surprised to witness a talk wherein Perry presented the N-localizer without attributing its origin to Brown [57]. When Perry, *et al.* subsequently proposed the Pfizer image-guided stereotactic system [6], which comprised N-localizers instead of slotted plates, they cited one [1] of Brown's original articles that had introduced the N-localizer one year earlier [1-2]. Several months before Perry, *et al.* proposed the Pfizer image-guided stereotactic system, Brown, *et al.* had already proposed the Brown-Roberts-Wells (BRW) image-guided stereotactic system [5].

The efforts of Perry, *et al.* to adapt an early-generation Leksell frame for CT imaging by attaching three slotted plates to that frame were unsuccessful [6, 26]. Perry, *et al.* abandoned the slotted plate, adopted instead the N-localizer, and never published a description of the slotted plate attached to a stereotactic frame.

However, Perry himself described three slotted plates attached to a stereotactic frame in his application to the U.S. Patent and Trademark Office dated April 13, 1979. The resulting patent issued on July 27, 1982 [58], and was the first public disclosure of Perry's technique of attaching three slotted plates to a stereotactic frame. Prior to that first public disclosure, Perry had disclosed privately to Brown, *circa* January 1979, three slotted plates attached to a stereotactic frame. An entry in Brown's notebook recounts his phone conversation with Perry that occurred *circa* January 1979. That entry includes a drawing of the slotted plate and reports that "John Perry of Pfizer began working on a localizing system, according to him in the fall of 1978. This system, as I understand it, consisted of 3 plates having vertical grooves in them" [57]. Brown's drawing and report, which are dated October 14, 1979, prove his awareness of Perry's slotted-plate technique three years prior to the first public disclosure of that technique and hence corroborate Brown's account of his phone conversation with Perry.

Perry's earliest description of the N-localizer was cursory and limited to only two sentences in his patent that devoted detailed explanations and five drawings to a thorough description of his slotted-plate technique [58]. When challenged by Brown via a Patent Interference proceeding before the U.S. Patent and Trademark Office, Perry failed to provide any evidence whatsoever of having invented the N-localizer. Consequently, Perry conceded "priority of invention" to Brown [59] and the U.S. Patent and Trademark Office awarded patent protection for the N-localizer to Brown [60]. The documents [1, 47-48, 54-56, 59] that the U.S. Patent and Trademark Office examined prior to awarding patent protection to Brown instead of Perry are a matter of public record. Those documents may be obtained from the U.S. Patent and Trademark Office by requesting a copy of the folder for Patent Interference No. 101267. In order to facilitate access to those documents, copies are included in the Appendices to this article.

## Conclusions

Brown invented the N-localizer and built the first CT-compatible stereotactic frame in 1978. The N-localizer has become an important tool for modern neurosurgery and has achieved widespread use in image-guided stereotactic neurosurgery and radiosurgery. Beginning in 1979, six different stereotactic frames incorporated the N-localizer. For each frame, the inclusion of the N-localizer was derivative and originated from Brown's prior research.

## Appendices



January 15, 1979

Dr. Dade Lunsford  
Dr. Arthur Rosenbaum  
Dr. David Jorvik  
University of Pittsburgh  
School of Medicine  
Pittsburgh, Pennsylvania 15261

Gentlemen:

I hope in our brief time together on Sunday you felt we accomplished your goals. Here is my description of the technique.

Once the fiducial plate file is set up for the frame, I think the clinical sequence of events will go something like this:

1. Mount the frame to the patient's head.
2. Attach the fiducial plates to the frame.
3. Scan the patient at the narrowest slice thickness until the target is found in the scan (this must be accomplished in a slice which intersects each fiducial plate at least one slot from each end, eg. more than one and less than the total number of slots must be seen in each of the three fiducial plates).
4. Fill out the Stereotaxic Data Sheet:  
Note the image scale factor:  
For each fiducial:  
a. Count the number of slots seen.  
b. Using the cursor, measure the CT scan coordinates of the slot which ends in the plane of the slice.  
Then stellarily measure the CT scan coordinates of the target.
5. Run the TAXIF program, typing in the requested data and recording the calculated frame coordinates of the target on the data sheet.
6. Check the comparisons of the lengths of vectors A, B, and C. We know that 1 mm accuracy can be obtained with differences of up to 0.35 cm. A large difference indicates something is wrong.
7. Remove the fiducial plates.
8. Carry out the classical stereotaxic procedure to be assured that the computer-calculated coordinates are consistent with the anatomy.
9. Set the calculated z coordinate into the probe carrier.
10. Set the side bars at the calculated Y coordinate.
11. Attach the probe carrier to the side bars at the calculated X coordinate. At this point the target is exactly centered at the origin of the frame's spherical coordinate system. Any angle (theta or phi) may be selected for the approach.

9052 Old Annapolis Road Columbia, Maryland 21045

EXHIBIT B

-2-

12. Check that the depth indicator on the probe carrier is set at zero.
13. Check that the probe length stop is correct using the length fixture.
14. Insert the probe.

In time, I hope to eliminate the transcribing of the coordinates and the rekeying of them into the TAXIF program.

The critically important points should be made about the procedure. First, the CT scan resolution element is large enough to cause errors of more than 1 mm, therefore, great care must be taken in reading CT scan coordinates and in counting fiducial bars. There is very little the computer can do to check the accuracy of its input. Second, the slice thickness in even the best circumstances will substantially contribute to the error unless care is taken to center the target in the plane of the slice. I think one of the more subtle contributions of the radiologist in this procedure will be his use of the partial volume effect to assure that the target is centered in the slice thickness, minimizing this potentially significant error. In this same vein, I would suggest that in any procedure, care be taken to aim for the center of the target volume.

The mathematics of our technique can be described as follows. The basic problem is to measure the coordinate transform between the coordinate system of the stereotaxic frame and the CT scanner. With this transform, the CT coordinates of a target can be converted to the corresponding frame coordinates.

To me, the innovation in our method is the technique for getting all the information on the transform and the target from one scan, thus eliminating the errors. The transform information is provided by three fiducial plates. With three non-collinear points measured in both coordinate systems, the transform can be calculated. The plates have a sequence of parallel slots of varying length such that by counting the slots which appear in the CT image one knows which slot ends in the plane of the slice. The plates are designed so that the frame coordinates of the end of each slot are known. Thus, counting the number of slots provides the frame coordinates of the fiducial. The CT scan coordinates of the fiducials are measured from the CT image.

There are several ways to actually calculate the transform. I approach it as a physicist might.

-3-

Let capital letters denote frame coordinates and lower case letters denote CT scan coordinates, and let:

$$\begin{aligned} F_1, F_2 &= \text{fiducial 1} \\ F_3, F_4 &= \text{fiducial 2} \\ F_5, F_6 &= \text{fiducial 3} \\ T, t &= \text{target} \end{aligned}$$

Define:

$$\begin{aligned} \vec{A} &= F_2 - F_1 & \vec{B} &= F_4 - F_1 \\ \vec{B} &= F_4 - F_1 & \vec{C} &= F_6 - F_1 \\ \vec{X} &= T - F_1 & \vec{Z} &= t - F_1 \end{aligned}$$

The problem is to find  $\vec{T}$ , the target frame coordinates.

Since  $\vec{A}$ ,  $\vec{B}$ , and  $\vec{Z}$  lie in the same plane and  $\vec{A}$  and  $\vec{B}$  are not collinear, we can find by solution of the simultaneous equations,  $\alpha$  and  $\beta$  such that:

$$\vec{Z} = \vec{t} - F_1 = \alpha \vec{A} + \beta \vec{B}$$

The equivalent in frame coordinates is:

$$\vec{X} = T - F_1 = \alpha \vec{A} + \beta \vec{B}$$

Thus:

$$\vec{T} = F_1 + \alpha \vec{A} + \beta \vec{B}$$

As you can see, the problem is trivial, mathematically. If it turns out to be medically useful, I think this will be another strong statement on behalf of CT. I look forward to visiting Pittsburgh in the very near future.

Sincerely,


John H. Perry  
Director  
Research and Development

JHP/cw

FIGURE 4: Appendix 1: John Perry Letter, pp. 1-3, January 15,



1979



**FIGURE 5: Appendix 2: Richard Matthews Letter, pp. 1-7,**

January 26, 1979

---




FIGURE 6: Appendix 3: Russell Brown Notebook 1, pp. 26-30,

May 24, 1978

---

80: Frame built should have 3 diagonals equally spaced around its perimeter:



There would thus be a rod at each 60° increment around the frame. This arrangement would provide the greatest accuracy in use of diagonals.

2/17/78

8/28/78

I have finished the necessary programming to allow simulated surgery. The method described on pages 75-80 of this notebook showing how to use the diagonals to map each scan slice into the frame coordinate system works beautifully. The reason that

I can say this is as follows:

I draw the rod and diagonal contours on the picture system as they are mapped into the frame coordinate system. I also draw an outline of the frame model as a collection of 2 circles, 2 arches, rods and diagonals, L shapes representing the displacement of the spheres lateral from the middle of the arches, a dashed line indicating the direction of probe insertion, and a probe. This is illustrated below:



82: The diagonal lines pass through the center of each ellipse along the diagonal lines. This indicates that the scan slices are correctly mapped into the frame coordinate system.

I have been placing the tip of the simulated probe at the edge of the various lucite sphere contours and recording the angle settings on the simulated frame as well as the depth of probe insertion (angles to nearest 1/10°, probe insertion to nearest mm.)

I then apply these settings to the frame and pass a lucite rod as a probe. In all cases so far the tip of the rod is within 2 mm from being "on target", that is, it is usually below and to the left of the intended point of contact

with the sphere. The depth by which the probe is actually inserted is within 2 mm of the predicted insertion depth. I think these deviations from ideal appear to have a pattern in lieu of being random. This would indicate that the frame is warped but that it has negligible play.

I intend to calibrate the frame by trying to hit each sphere using the predicted settings and insertion depth, and by comparing those settings with the settings and depth actually required to hit the spheres "dead center". If a constant pattern emerges I will have found the correct calibration factors to add to the predicted frame settings when I apply these settings to the actual frame.

5/20/78

FIGURE 7: Appendix 4: Russell Brown Notebook 1, pp. 80-83,

August 28, 1978

54

§ The square root of the sums of the squares of the three individual errors is taken to be the error for a probe placement. This is possible because these 3 errors are approximately orthogonal to one another. For the 20 probe placements documented on the preceding page the mean error is 2.03 mm and the standard deviation is 0.47 mm.

I was present for a number of these tests, witnessed same, and loaned my hand calculator for calculation of the mean and sigma on 1/20/79

1/20/79

1/25/79

I have spoken with Brian Heightman (sp?) of Pfizer Medical, Inc. about my Stereotactic project. He says that John Perry of Pfizer is working on a similar project and has found that the Pfizer reconstruction algorithm is unaffected (or minimally affected) by aluminum. I would, of course, prefer to use aluminum for the frame fabrication because it is more durable than plastic. 1/25/79

**FIGURE 8: Appendix 5: Russell Brown Notebook 2, p. 54, January 25, 1979**



10/14/79

99

I anticipate some trouble over my patent claims, both the localizing rod system and the concept of a frame which allows passage of a probe to any point inside the frame from any direction through a hemisphere. The reasons that I anticipate trouble are as follows:

- 1) John Perry of Pfizer began working on a localizing system, according to him in the fall of 1978. This system, as I understand it, consisted of 3 plates having vertical grooves in them:



During a telephone conversation with John Perry (I think in January or February of 1979) I pointed out to John the merits of a simple diagonal rod bounded by 2 vertical rods. He

100

agreed that this was a better localizing system than plates with vertical slots. Then in May of 1979 (I believe it was May) at the ASNR meeting John presented a frame with such diagonal rods. He did not, however, acknowledge to his audience that I had advised him to use diagonal rods.

Since that time Art Rosenbaum has denied, once to Jim Nelson and once to Trent Wells, that I gave John Perry any ideas. He simply has said that John Perry was working on a localizing system before he spoke to me. This is true, but the system he was working on was the plate, not the diagonal rod.

- 2) Art Rosenbaum told Trent Wells last week at the CNS meeting in Las Vegas that he was involved in and at the point of building 20 frames of some design

but that after seeing the Brown-Roberts-Wells frame he could promise Trent that he would buy 20 Brown-Roberts-Wells frames instead. He (Rosenbaum) stated that the concept of passing a probe to any point inside the frame from any direction through a hemisphere was quite different than the frame he was planning to build. Apparently, from Trent's description of that frame, it allows probe insertion to a target point through a pyramidal set of pathways:



Trent says this type of geometry is like the old Borsely-Clark stereotaxic frame.

102

In addition, Rosenbaum was very interested in interchangeable localizing rod and arch systems which lock onto the head mounting ring in the same manner. His frame apparently does not have such interchangeable systems.

Rosenbaum took a few pictures of the ~~frame~~ Brown-Roberts-Wells frame. Trent reminded him that the frame is protected by patent claims.

Red, discussed  
and collected  
10/14/79

10/14/79

d

FIGURE 9: Appendix 6: Russell Brown Notebook 3, pp. 99-102,

October 14, 1979



PATENT  
RECEIVED  
DEC 6 1985  
BOARD OF APPEALS  
AND INTERFERENCES

IN THE UNITED STATES PATENT AND TRADEMARK OFFICE

Russell A. Brown,	)	
	)	
Junior Party,	)	Patent Interference
	)	
v.	)	No. 101,267
	)	
John H. Perry,	)	
	)	
Senior Party.	)	

CONCESSION OF PRIORITY

707 Wilshire Boulevard  
Los Angeles, California 90017

Commissioner of Patents  
and Trademarks  
Washington, D. C. 20231

Sir:

Based on an exchange of information herein, the undersigned hereby concedes priority with regard to the subject matter of this interference. Specifically, this constitutes a concession of priority with regard to the subject matter of Counts 1 through 18 in the interference.

Dated: Nov. 27, 1985

John H. Perry

Consent to concession of priority by Assignee:

Dated: \_\_\_\_\_

Russell A. Brown

Respectfully submitted,

Dated: Dec 3, 1985

B. G. Nilsson  
Registration No. 17,350

Docket No. 2568-101  
(213) 620-0600

**FIGURE 10: Appendix 7: John Perry Concession of Priority,  
November 27, 1985**

## Additional Information

## Disclosures

**Conflicts of interest:** In compliance with the ICMJE uniform disclosure form, all authors declare the following: **Payment/services info:** All authors have declared that no financial support was received from any organization for the submitted work. **Financial relationships:** All authors have declared that they have no financial relationships at present or within the previous three years with any organizations that might have an interest in the submitted work. **Other relationships:** All authors have declared that there are no other relationships or activities that could appear to have influenced the submitted work.

## Acknowledgements

The authors thank Barbara Pedrick of the Stanford University Department of Neurosurgery for assistance in obtaining hard-to-find journal articles.

## References

1. Brown RA: A computerized tomography-computer graphics approach to stereotaxic localization. *J Neurosurg.* 1979, 50:715–720. [10.3171/jns.1979.50.6.0715](#)
2. Brown RA: A stereotactic head frame for use with CT body scanners. *Invest Radiol.* 1979, 14:300–304. [10.1097/00004424-197907000-00006](#)
3. Brown RA, Nelson JA: Invention of the N-localizer for stereotactic neurosurgery and its use in the Brown-Roberts-Wells stereotactic frame. *Neurosurgery.* 2012, 70 (2 Suppl Operative):173–176. [10.1227/NEU.0b013e318246a4f7](#)
4. Brown RA, Roberts TS: A stereotaxic frame and computer software for use with CT body scanners. *Stereotactic Cerebral Irradiation: Proceedings of the INSERM Symposium on Stereotactic Irradiation.* Szikla G (ed): Elsevier/North-Holland Biomedical Press, Amsterdam; 1979. 13:25–27.
5. Brown RA, Roberts TS, Osborn AE: Stereotaxic frame and computer software for CT-directed neurosurgical localization. *Invest Radiol.* 1980, 15:308–312. [10.1097/00004424-198007000-00006](#)
6. Perry JH, Rosenbaum AE, Lunsford LD, Swink CA, Zorub DS: Computed tomography-guided stereotactic surgery: conception and development of a new stereotactic methodology. *Neurosurgery.* 1980, 7:376–381. [10.1097/00006123-198010000-00011](#)
7. Goerss S, Kelly PJ, Kall B, Alker GJ Jr: A computed tomographic stereotactic adaptation system. *Neurosurgery.* 1982, 10:375–379. [10.1097/00006123-198203000-00014](#)
8. Dubois PJ, Nashold BS, Perry J, Burger P, Bowyer K, Heinz ER, Drayer BP, Bigner S, Higgins AC: CT-guided stereotaxis using a modified conventional stereotaxic frame. *AJNR Am J Neuroradiol.* 1982, 3:345–351.
9. Leksell L, Leksell D, Schwebel J: Stereotaxis and nuclear magnetic resonance. *J Neurol Neurosurg Psychiatry.* 1985, 48:14–18. [10.1136/jnnp.48.1.14](#)
10. Couldwell WT, Apuzzo ML: Initial experience related to the use of the Cosman-Roberts-Wells stereotactic instrument. Technical note. *J Neurosurg.* 1990, 72:145–148. [10.3171/jns.1990.72.1.0145](#)
11. Lunsford LD: A dedicated CT system for the stereotactic operating room. *Appl Neurophysiol.* 1982, 45:374–378. [10.1159/000101625](#)
12. Apuzzo ML, Sabshin JK: Computed tomographic guidance stereotaxis in the management of intracranial mass lesions. *Neurosurgery.* 1983, 12:277–285. [10.1227/00006123-198303000-00005](#)
13. Alker G, Kelly PJ, Kall B, Goerss S: Stereotaxic laser ablation of intracranial lesions. *AJNR Am J Neuroradiol.* 1983, 4:727–730.
14. Heilbrun MP, Roberts TS, Apuzzo ML, Wells TH Jr, Sabshin JK: Preliminary experience with Brown-Roberts-Wells (BRW) computerized tomography stereotaxic guidance system. *J Neurosurg.* 1983, 59:217–222. [10.3171/jns.1983.59.2.0217](#)
15. Leksell L: Stereotactic radiosurgery. *J Neurol Neurosurg Psychiatry.* 1983, 46:797–803. [10.1136/jnnp.46.9.797](#)
16. Thomas DG, Anderson RE, du Boulay GH: CT-guided stereotactic neurosurgery: experience in

- 24 cases with a new stereotactic system. *J Neurol Neurosurg Psychiatry*. 1984, 47:9–16. [10.1136/jnnp.47.1.9](https://doi.org/10.1136/jnnp.47.1.9)
17. Bullard DE, Nashold BS Jr, Osborn D, Burger PC, Dubois P: CT-guided stereotactic biopsies using a modified frame and Gildenberg techniques. *J Neurol Neurosurg Psychiatry*. 1984, 47:590–595. [10.1136/jnnp.47.6.590](https://doi.org/10.1136/jnnp.47.6.590)
18. Apuzzo ML, Chandrasoma PT, Zelman V, Giannotta SL, Weiss MH: Computed tomographic guidance stereotaxis in the management of lesions of the third ventricular region. *Neurosurgery*. 1984, 15:502–508. [10.1227/00006123-198410000-00005](https://doi.org/10.1227/00006123-198410000-00005)
19. Thomas DG, Davis CH, Ingram S, Olney JS, Bydder GM, Young IR: Stereotaxic biopsy of the brain under MR imaging control. *AJNR Am J Neuroradiol*. 1986, 7:161–163.
20. Lunsford LD, Martinez AJ, Latchaw RE: Stereotaxic surgery with a magnetic resonance- and computerized tomography-compatible system. *J Neurosurg*. 1986, 64:872–878. [10.3171/jns.1986.64.6.0872](https://doi.org/10.3171/jns.1986.64.6.0872)
21. Heilbrun MP, Sunderland PM, McDonald PR, Wells TH Jr, Cosman E, Ganz E: Brown-Roberts-Wells stereotactic frame modifications to accomplish magnetic resonance imaging guidance in three planes. *Appl Neurophysiol*. 1987, 50:143–152. [10.1159/000100700](https://doi.org/10.1159/000100700)
22. Apuzzo ML, Petrovich Z, Luxton G, Jepson JH, Cohen D, Breeze RE: Interstitial radiobrachytherapy of malignant cerebral neoplasms: rationale, methodology, prospects. *Neurol Res*. 1987, 9:91–100.
23. Apuzzo ML, Chandrasoma PT, Cohen D, Zee CS, Zelman V: Computed imaging stereotaxy: experience and perspective related to 500 procedures applied to brain masses. *Neurosurgery*. 1987, 20:930–937. [10.1227/00006123-198706000-00019](https://doi.org/10.1227/00006123-198706000-00019)
24. Debaene A, Gomez A, Lavieille J, Alessandri C, Legre J: Stereotactic CT localization and biopsy of brain tumours using the Leksell frame. A study of 45 cases. *J Neuroradiol*. 1988, 15:266–275.
25. Loeffler JS, Alexander E 3rd, Siddon RL, Saunders WM, Coleman CN, Winston KR: Stereotactic radiosurgery for intracranial arteriovenous malformations using a standard linear accelerator. *Int J Radiat Oncol Biol Phys*. 1989, 17:673–677. [10.1016/0360-3016\(89\)90123-5](https://doi.org/10.1016/0360-3016(89)90123-5)
26. Lunsford LD, Coffey RJ, Cojocaru T, Leksell D: Image-guided stereotactic surgery: a 10-year evolutionary experience. *Stereotact Funct Neurosurg*. 1990, 54-55:375–387. [10.1159/000100239](https://doi.org/10.1159/000100239)
27. Andoh K, Nakamae H, Ohkoshi T, Odagiri K, Kyuma Y, Hayashi A: Technical note: enhanced MR-guided stereotaxic brain surgery with the patient under general anesthesia. *AJNR Am J Neuroradiol*. 1991, 12:135–138.
28. Pell MF, Thomas DG: The initial experience with the Cosman-Roberts-Wells stereotactic system. *Br J Neurosurg*. 1991, 5:123–128. [10.3109/02688699108998457](https://doi.org/10.3109/02688699108998457)
29. Martinez R, Vaquero J: Image-directed functional neurosurgery with the Cosman-Roberts-Wells stereotactic instrument. *Acta Neurochir (Wein)*. 1991, 113:176–179. [10.1007/BF01403204](https://doi.org/10.1007/BF01403204)
30. Pillay PK, Barnett G, Awad I: MRI-guided stereotactic placement of depth electrodes in temporal lobe epilepsy. *Br J Neurosurg*. 1992, 6:47–53. [10.3109/02688699209002901](https://doi.org/10.3109/02688699209002901)
31. Maciunas RJ, Kessler RM, Maurer C, Mandava V, Watt G, Smith G: Positron emission tomography imaging-directed stereotactic neurosurgery. *Stereotact Funct Neurosurg*. 1992, 58:134–140. [10.1159/000098986](https://doi.org/10.1159/000098986)
32. Kondziolka D, Dempsey PK, Lunsford LD, Kestle JR, Dolan EJ, Kanal E, Tasker RR: A comparison between magnetic resonance imaging and computed tomography for stereotactic coordinate determination. *Neurosurgery*. 1992, 30:402–406. [10.1227/00006123-199203000-00015](https://doi.org/10.1227/00006123-199203000-00015)
33. Sofat A, Hughes S, Briggs J, Beaney RP, Thomas DG: Stereotactic brachytherapy for malignant glioma using a relocatable frame. *Br J Neurosurg*. 1992, 6:543–548. [10.3109/02688699209002371](https://doi.org/10.3109/02688699209002371)
34. Liang JA, Lin FJ, Tsai MD, Tu CP, Hsiao AC: Implementation of stereotactic focal radiotherapy for intracranial arteriovenous malformations using a linear accelerator. *J Formos Med Assoc*. 1993, 92:988–994.
35. Kratimenos GP, Thomas DG, Shorvon SD, Fish DR: Stereotactic insertion of intracerebral electrodes in the investigation of epilepsy. *Br J Neurosurg*. 1993, 7:45–52. [10.3109/02688699308995055](https://doi.org/10.3109/02688699308995055)
36. Pell MF, Thomas DG, Kratimenos GP: Stereotactic management of intrinsic brain stem lesions. *Ann Acad Med Singapore*. 1993, 22:447–451.

37. Matsumoto K, Higashi H, Tomita S, Ohmoto T: Pineal region tumours treated with interstitial brachytherapy with low activity sources (192-iridium). *Acta Neurochir (Wien)*. 1995, 136:21–28. [10.1007/BF01411431](https://doi.org/10.1007/BF01411431)
38. Dormont D, Cornu P, Pidoux B, Bonnet AM, Biondi A, Oppenheim C, Hasboun D, Damier P, Cuchet E, Philippon J, Agid Y, Marsault C: Chronic thalamic stimulation with three-dimensional MR stereotactic guidance. *AJNR Am J Neuroradiol*. 1997, 18:1093–1107.
39. Levivier M, Massager N, Wikler D, Lorenzoni J, Ruiz S, Devriendt D, David P, Desmedt F, Simon S, Van Houtte P, Brotchi J, Goldman S: Use of stereotactic PET images in dosimetry planning of radiosurgery for brain tumors: clinical experience and proposed classification. *J Nucl Med*. 2004, 45:1146–1154.
40. Lunsford LD, Niranjan A, Khan AA, Kondziolka D: Establishing a benchmark for complications using frame-based stereotactic surgery. *Stereotact Funct Neurosurg*. 2008, 86:278–287. [10.1159/000147636](https://doi.org/10.1159/000147636)
41. Kondziolka DS, Lunsford LD: CT in image-guided surgery. *Textbook of Stereotactic and Functional Neurosurgery*. Lozano AM, Gildenberg PL, Tasker RR (ed): Springer Verlag, New York; 2009. 1:619-629. [10.1007/978-3-540-69960-6\\_40](https://doi.org/10.1007/978-3-540-69960-6_40)
42. Rosenbaum AE, Lunsford LD, Perry JH: Computerized tomography guided stereotaxis: A new approach. *Appl Neurophysiol*. 1980, 43:172-173. [10.1159/000102252](https://doi.org/10.1159/000102252)
43. Leksell L, Jernberg B: Stereotaxis and tomography. A technical note. *Acta Neurochir (Wien)*. 1980, 52:1–7. [10.1007/BF01400939](https://doi.org/10.1007/BF01400939)
44. Roberts TS, Brown R: Technical and clinical aspects of CT-directed stereotaxis. *Appl Neurophysiol*. 1980, 43:170-171. [10.1159/000102251](https://doi.org/10.1159/000102251)
45. Lunsford LD, Rosenbaum AE, Perry J: Stereotactic surgery using the “therapeutic” CT scanner. *Surg Neurol*. 1982, 18:116-122. [10.1016/0090-3019\(82\)90369-X](https://doi.org/10.1016/0090-3019(82)90369-X)
46. Lunsford LD, Niranjan A, Kassam A, Khan A, Amin D, Kondziolka D: Intraoperative imaging: evolutions, options, and practical applications. *Clin Neurosurg*. 2008, 55:76–86.
47. Perry JH: Letter to Lunsford D, Rosenbaum A, and Zorub D. Available in Appendix 1 and U.S. Patent and Trademark Office Patent Interference folder 101267. 1979, 1-3.
48. Matthews RS: Letter to Nixon LS. Available in Appendix 2 and U.S. Patent and Trademark Office Patent Interference folder 101267. 1979, 1-7.
49. Lunsford LD, Leksell L, Jernberg B: Probe holder for stereotactic surgery in the CT scanner: a technical note. *Acta Neurochir (Wien)*. 1983, 69:297-304. [10.1007/BF01401817](https://doi.org/10.1007/BF01401817)
50. Lee SH, Villafana T, Lapayowker MS: CT intracranial localization with a new marker system. *Neuroradiology*. 1978, 16:570-571. [10.1007/BF00395364](https://doi.org/10.1007/BF00395364)
51. Villafana T, Lee SH, Lapayowker MS: A device to indicate anatomical level in computed tomography. *J Comput Assist Tomogr*. 1978, 2:368-371. [10.1097/00004728-197807000-00028](https://doi.org/10.1097/00004728-197807000-00028)
52. Dohrmann GJ, Geehr RB, Robinson F, Allen WE 3rd, Orphanoudakis, SC: Small hemorrhages vs. small calcifications in brain tumors: difficulty in differentiation by computed tomography. *Surg Neurol*. 1978, 10:309-312.
53. Schultz E, Felix R: Phantom measurements of spatial resolution and the partial-volume-effect in computer tomography (article in German). *Rofo*. 1978, 129:673-678. [10.1055/s-0029-1231185](https://doi.org/10.1055/s-0029-1231185)
54. Brown RA: Notebook 1. Available in Appendix 3 and U.S. Patent and Trademark Office Patent Interference folder 101267. 1978, 26-30.
55. Brown RA: Notebook 1. Available in Appendix 4 and U.S. Patent and Trademark Office Patent Interference folder 101267. 1978, 80-83.
56. Brown RA: Notebook 2. Available in Appendix 5 and U.S. Patent and Trademark Office Patent Interference folder 101267. 1979, 54.
57. Brown RA: Notebook 3. Available in Appendix 6. 1979, 99-102.
58. Perry JH: Stereotactic surgery apparatus and method. U.S. patent 4341220. 1982, 1-10. Accessed: September 27, 2013: <https://www.google.com/patents/US4341220>.
59. Perry JH: Concession of Priority. Available in Appendix 7 and U.S. Patent and Trademark Office Patent Interference folder 101267. 1985, 1.
60. Brown RA: System using computed tomography as for selective body treatment. U.S. patent 4608977. 1986, 1-12. Accessed: September 27, 2013: <https://www.google.com/patents/US4608977>.